



P 937.382.1451  
F 937.382.1191

70 Weil Way  
Wilmington, Ohio 45177

877.733.0911  
[www.ferno.com](http://www.ferno.com)

2012

## DECLARATION OF CONFORMITY

Manufacturer: Ferno-Washington, Inc.

Manufacturer's Address: 70 Weil Way  
Wilmington, Ohio  
USA

Name of Device(s): Traverse Rescue Stretcher (TRS)

Class of Device(s): I

Intended Use: Patient Transport

In accordance with European Community Council Directive 93/42/EEC, Annex VII, Ferno-Washington Inc. declares the products that bear the CE mark conform with all the provisions of this Directive that apply to them.

Ferno-Washington Inc. products have been determined to be a Class I device according to Article 9.1 and using classification criteria from Annex IX of the Directive.

Ferno-Washington Inc. or representative will make available upon request all applicable technical documentation to allow assessment of conformity of its products in accordance with Annex VII, Sections 2 and 3.

Ferno-Washington Inc. has a systematic procedure in place to review experience gained from its devices in the post-production phase, taken appropriate corrective action based on this experience and report malfunctions to the competent authorities in accordance with Annex VII, Section 4.

Thomas E. Livingston  
Director of Quality